

Certificate of analysis

Product Name/Description: TaKaRa PCR Human Papillomavirus Typing Set

Cat. #: 6603

Lot #: Specified on product label

Storage Conditions: -20 degrees C

Expiration Date: Specified on product label or package

Package Size: 100 Rxns

Package Contents:	1. HPVpU-1M Primer	25 pmol/μl	50 μl
	2. HPVpU-31B Primer	25 pmol/μl	50 μl
	3. HPVpU-2R Primer	25 pmol/μl	100 μl
	4. Control Template HPV-TB	1 ng/μl	50 μl
	5. Control Template HPV-TM	1 ng/μl	50 μl

Quality Control Data:

PCR test:

1) Expected product (63 bp) was obtained by PCR using Control Template HPV-TM supplied in the set and the primer set (HPVpU-1M and HPVpU-2R).

2) Expected product (61 bp) was obtained by PCR using Control Template HPV-TB supplied in the set and the primer set (HPVpU-31B and HPVpU-2R).

3) No specific amplification was confirmed by PCR using either of the primer sets and negative control.

It is certified that this product meets the above specifications, as reviewed and approved by the Quality Department of Takara Bio Inc.

Notice To Purchaser:

Please refer to the website.

Product Documents, Patent Markings, Trademarks, & Safety Information:

Refer to website for availability to download or contact your local office:
www.takara-bio.co.jp (Japanese)
www.takarabio.com (English)

Licence Statements:

Please visit our website and click the "license" button on the product's web page.

©2019 Takara Bio Inc. All Rights Reserved.

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions.